


KERN 440-45N

The classic balance in the laboratory

KERN



Construction

| | |
|-----------------------------------|-----------------|
| Dimension housing (W×D×H) | 165×230×80 mm |
| Dimensions weighing surface (W×D) | 130×130 mm |
| Material housing | plastic |
| Material weighing plate | stainless steel |

Functions

| | |
|--|-------------------|
| Number of keys for operation | 5 |
| Auto-off interval(s) in battery mode/rechargeable battery mode | 3 min off |
| Tare function | manual (multi) |
| Percentage determination | ✓ |
| Counting function | ✓ |
| Function for averaging under unstable weighing conditions | ✓ |
| Recipe/summing level | Rez A (Net total) |
| Underfloor weighing | Hook (optional) |
| Interfaces | RS-232 standard |

Counting

| | |
|---|-------------------|
| Smallest piece weight when piece counting - laboratory conditions | 200 mg |
| Smallest piece weight when piece counting - normal conditions | 2 g |
| Reference quantity | 5, 10, 20, 25, 50 |
| Counting resolution (laboratory conditions) | 5.000 |

Power Supply

| | |
|---|-------------------------------|
| Input voltage power supply / power [Max] | 100 - 240 V AC 50/60 Hz 0.5 A |
| Input voltage device / power [Max] | 9 V, 300 mA |
| Supplied plug-in power supply | Power supply unit |
| Plug-in power supply / adapter for countries - included with the delivery | EURO UK US CH |
| Plug-in power supply / adapter for countries - optional | EURO UK US CH |
| Rechargeable battery optional | Rchrg. battery optional |
| Rechargeable battery charging time | 10 h |
| Rechargeable battery operating time - backlight off | 20 h |
| Battery | 9 V block |
| Battery operating time | 20 h |

Category

| | |
|------------------|---------------------|
| Brand | KERN |
| Product category | Laboratory balances |
| Product group | Precision balances |

Measuring System

| | |
|--|--|
| Readability [d] | 0,1 g |
| Weighing capacity [Max] | 1000 g |
| Eccentric loading at 1/3 [Max] | 0,3 g |
| USP Minimum weight of sample (k = 2, U = 0.1%) | 200 g |
| Reproducibility | 0,1 g |
| Linearity | ± 0,2 g |
| Weighing system | Strain gauge |
| Resolution | 10.000 |
| Units | kg g dwt tl (Tw) tl (HK) ozt mo lb oz ffa tol |
| Default unit | g |
| Stabilization time | 3 s |
| Warm-up time | 30 min |
| Adjustment options | Adjusting with external weight |
| Recommended adjusting weight | 1 kg (M1) |
| Possible calibration points | 200 g; 500 g; 700 g; 1 kg |

Approval

| | |
|---------|---|
| CE mark | ✓ |
|---------|---|

Display

| | |
|----------------------|-------|
| Display digit height | 15 mm |
| Display backlight | ✓ |

KERN 440-45N



The classic balance in the laboratory

Environmental conditions

| | |
|-------------------------------|-------|
| Humidity of environment [Max] | 80 % |
| Ambient temperature [Min] | 5 °C |
| Ambient temperature [Max] | 35 °C |

Packing & Shipping


| | |
|------------------------------|----------------|
| Readability force [d] (N) | 1 d |
| Shipping method | Parcel service |
| Dimensions packaging (W×D×H) | 210×350×170 mm |
| Net weight approx. | 1,2 kg |
| Gross weight approx. | 1,6 kg |

Services (optional)

| | |
|---|---------|
| Article number for adjustment at the location of installation | 961-247 |
| Article number for DAkkS calibration | 963-127 |
| Article number for certificate of conformity (verification) | 969-517 |

Piktogramme

STANDARD



OPTION

